Electronic Development Labs, Inc.
The Temperature People®
Since 1943

CERTIFICATE OF GENERAL CALIBRATION

FOR DIGITAL INSTRUMENTS

Model Number

Serial Number

This certificate is valid for one year from date of purchase.

This instrument comes factory calibrated to NIST standards and has been checked at the following points: 0 °C, 37.78 °C (100 °F), 537.78 °C (1000 °F), 1037.78 °C (1900 °F) ±1°.

Under normal operating conditions, we recommend yearly calibration.

Electronic Development Laboratories, Inc. (EDL) certifies this instrument was calibrated in accordance with applicable calibration procedures during the instrument manufacturing process.

Electronic Development Laboratories, Inc. (EDL) further certifies this instrument meets or exceeds published measurement specifications (unless otherwise noted) and has been calibrated using standards traceable to the National Institute of Standards and Technology (NIST) or National Measurement Laboratory (NML/ROC).

Quality Assurance Manager

Electronic Development Labs, Inc.
244 Oakland Drive • Danville, VA 24540
www.edl-inc.com • 434.799.0807